

KwikSwitch Accessories

Perfect add-ons for KwikSwitch product series thermocouple instruments

Get more use out of your folding thermocouple instrument with the 9418 Rubber Boot and 53337-K Replacement DuraNeedle Probe.

Compatible with the 94100 KwikSwitch, 94100-01 Steak Genius, and 94100-B KwikSwitch BLE, these accessories are sold separately.

9418 Rubber Boot

Helps protect any KwikSwitch thermocouple instrument with an extra layer of durable rubber.

Includes:

- Magnets to secure the instrument to any magnetic metal surface
- A lanyard to help prevent accidental drops during use

53337-K Replacement DuraNeedle Probe

Securely folds into any KwikSwitch thermocouple instrument and maintains the total system accuracy of the instrument when used to replace the included probe.

The probe features a screw-in flanged connection to prevent the needle from getting stuck in dense foods.



Specifications

53337-K Replacement DuraNeedle Probe

- **Temperature Range:** -40° to 500°F (-40° to 260°C)
- **Probe Accuracy:** In-field replacement will maintain $\pm 1^\circ\text{F}$ Total System Accuracy
- **Response Time:** 1 second (liquid)
- **Probe Length:** 3.75" (95 mm)
- **Probe Diameter:** 0.13" (3.3 mm) with 0.067" (1.70 mm) reduced tip
- **Probe storage slot designed to fit 53337-K probe**
- **Warranty:** Limited 1-year probe warranty

Contact your Cooper-Atkins representative for more information.



Copeland Cold Chain LP
67-1892 | V0125

To learn more, visit copeland.com/cooper-atkins

©2025 Copeland LP.

COOPER-ATKINS